

ALLINA HEALTH LABORATORY

Memo

To: Allina Health Laboratory outreach clients

From: Allina Health Laboratory outreach services

Date: May 28, 2020

Re: Greiner Vacuette® 2 mL, 9NC Coagulation Sodium Citrate 3.2% tube recall

Allina Health Laboratory has received notice of a recall on the Greiner Vacuette® 2 mL, 9NC Coagulation Sodium Citrate 3.2% tubes due to an incorrect amount of citrate added to some of the tubes during manufacturing.

Product: Greiner Vacuette® 2 mL, 9NC Coagulation Sodium Citrate 3.2% Tubes

Lot #: B1908355

Exp: 200809

These tubes are included as a component of the Aspirin therapy and P2Y12 collection kit/tubes provided by Allina Health Laboratory.



Please check all kits that you may have in stock for tubes affected by this recall, and immediately remove them from use. Return any kits containing the recalled tubes to Allina Health Laboratory along with a completed [Supply Return form](#).

If you have any questions, contact your account representative for assistance.