ALLINA HEALTH LABORATORY

To: Allina Health Laboratory outreach clients

From: Allina Health Laboratory outreach services

Date: October 13, 2023

Re: aPTT Test Change

APTT

490/85730.0

Quick Reference

Changing to a new lot of aPTT reagent requires thorough validation testing to establish new Normal Reference and Therapeutic Heparin Response ranges. This process occurs approximately every 12 months when a new lot of reagent is made available by the instrumentation reagent vendor.

Details

As we prepare for changing to a new aPTT reagent lot Allina Health Laboratories Coagulation Department undertakes a new "normal range" study using 25 specimens from persons known to be free of coagulation abnormalities and not on any current medications. These samples are run at every site, on each analyzer to give us a large volume dataset and a thorough representation of instrument comparability.

Heparin responsiveness of the new lot of aPTT reagent is also assessed in more than 100 patients currently on heparin therapy which is correlated with anti-Xa chromogenic heparin levels. This correlation study is the basis for the new heparin aPTT therapeutic range.

New Ranges effective October 17, 2023:

	Current (seconds)	New (seconds)
Normal Reference Range	24 - 33	25 - 33
Heparin Therapeutic Range	70 - 95	78 - 109