

ALLINA HEALTH LABORATORY

Memo

To: Allina Health Laboratory outreach clients

From: Allina Health Laboratory outreach services

Date: October 13, 2021

Re: COVID collection kit supply issue


Allina Health Laboratory has received notification that some of the COVID-19 collection kits that we recently received from our supplier had incorrect swabs included. The kits were labelled as NP (nasopharyngeal) kits, however, the swab contained in the kits were for OP collections.

TO DO:

If your site submits COVID-19 testing to Allina Health Laboratory for testing:

- Review the inventory of any collection kits that you may have received *since October 6, 2021*
 - **Correct** COVID Collection kits include a FLOQswabs NP swab
 - **Incorrect** COVID collection kits include a CLASSIQSwabs OP swab
- Return any incorrect kits back to Central Lab along with a completed [Supply Return Form](#)

Correct kit



Incorrect kit

