

# ALLINA HEALTH LABORATORY

## Memo

---

**To:** Allina Health Laboratory outreach clients

**From:** Allina Health Laboratory outreach services

**Date:** September 15, 2021

**Re:** Aspirin therapy and P2Y12 collection kit/tubes tube recall

---

Allina Health Laboratory has received notice of a recall of two lot numbers of the Greiner Bio-One Vacuette 2mL Sodium citrate tubes due to low draw volume and resulting clotting of the samples. These tubes are a component of the Aspirin therapy and P2Y12 collection kits.

The two lot numbers affected by this issue/recall are:

- Lot number B210533B, expiration 5/10/2022
- Lot number B210439J, expiration 4/13/82022

### What do you need to do?

- If you have Aspirin therapy and P2Y12 collections in stock at your site, examine the Sodium citrate tube(s) to see if any of the kits contain tubes with the affected lot numbers.
- If you have kits with the affected lot numbers, complete a [Supply return form](#) and send the kits, along with the completed form, back to the Central Laboratory to the attention of Diane Hutchinson
- Place an order in the [Supply catalog](#) for replacement kits
- If you have questions regarding the recall or return of affected supplies, contact your account representative or Client Services at 612-863-4678